

Instructions for Use

RealLine HSV-1 / HSV-2 Str-Format

ASSAY KIT FOR QUALITATIVE AND DIFFERENTIAL DETECTION OF DNA OF THE HERPES
SIMPLEX VIRUSES TYPE 1 AND TYPE 2 BY REAL TIME PCR METHOD










In vitro Diagnostics



RealLine HSV-1 / HSV-2 (Str-format)	VBD2195	96 tests
valid from	September 2019	

RealLine HSV-1 / HSV-2 Str-Format

Explanation of symbols used in labeling

	<i>In vitro</i> diagnostic medical device
	Batch code
	Catalogue number
	Contains sufficient for <n> tests
	Use-by-date
	Temperature limit
	Consult instructions for use
	Keep away from sunlight
	Manufacturer



BIORON Diagnostics GmbH

In den Rauhweiden 20
67354 Römerberg
Germany

Phone +49 6232 298 44 0

Fax: +49 6232 298 44 29

info@bioron.de

**RealLine HSV-1 / HSV-2
Str-Format**

Table of content:

1. INTENDED USE	4
2. KIT CONTENTS	5
3. PRINCIPLE OF THE METHOD	5
4. SPECIFICATIONS	6
5. PRODUCT USE LIMITATIONS	7
6. WARNING AND PRECAUTIONS	8
7. ADDITIONAL MATERIALS AND DEVICES REQUIRED BUT NOT SUPPLIED	8
8. PREPARATION OF THE SPECIMEN	9
9. PROCEDURE	10
10. DATA ANALYSIS AND INTERPRETATION	11
11. STORAGE AND TRANSPORTATION	12
ANNEX I: Settings for RealLine Cyclor and DT96:	12

RealLine HSV-1 / HSV-2 Str-Format

ASSAY KIT FOR QUALITATIVE AND DIFFERENTIAL DETECTION OF DNA OF THE HERPES SIMPLEX VIRUSES TYPE 1 AND TYPE 2 BY REAL TIME PCR METHOD

In vitro Diagnostics

1. INTENDED USE

Clinical information:

HSV-1 (Human Herpes Simplex Type 1) is an oral form of herpes, which typically affects the mouth and surrounding areas on the face. In addition HSV-1 can cause encephalitis, keratoconjunctivitis, stomatitis and skin infections. Transmission generally occurs through saliva and rarely through sex contact.

HSV-2 (Human Herpes Simplex Type 2) causes most cases of genital herpes as well as the neurological manifestations such as meningitis and radiculitis. Pregnant women and infants are risk groups for HSV infections, since when infants do contract neonatal herpes, they may suffer serious neurological damage, mental retardation or death. In addition HSV-2 is known to increase a person's risk of contracting HIV

RealLine HSV-1/HSV-2 (Str-format) assay kit is designed for detection of DNA and differential determination of herpes simplex virus types 1 and 2 isolated from clinical specimens using extraction kits

RealLine DNA-Express (REF VBC8899)

RealLine DNA-Extraction 2 (REF VBC8897)

RealLine DNA-Extraction 3 (REF VBC8889)

RealLine Extraction 100 (REF VBC8896)

RealLine HSV-1/HSV-2 (Str-format) kit is intended for differential detection of DNA of Herpes Simplex virus types 1 and 2 in clinical specimens (swabs of epithelial cells, tissue fluid, erosive-ulcerative skin lesion, mucosa, cerebrospinal fluid, saliva, semen, prostate fluid, urine, serum, plasma, whole blood, bronchoalveolar lavage) using the method of real-time polymerase chain reaction (PCR) with fluorescence detection of amplified product.

The Str-Format Kit contains 96 tubes (0.2ml) in strips with lyophilized Mastermix. 50 µl of extracted DNA have to be pipetted into the tube and the ready mastermix is diluted. The kit contains reagents required for 96 tests, including control samples and the positive control sample.

The kit is validated for use with: iQ™5 iCycler (Bio-Rad, USA). The kit is compatible with real-time PCR systems such as iQ™ iCycler, CFX™96 (Bio-Rad, USA). RealLine Cyclo (BIORON Diagnostics GmbH) and DT-96 (DNA-Technology, Russia).

RealLine HSV-1 / HSV-2 Str-Format

The use of:

- ! Extraction Kits for nucleic acids from clinical specimen from other supplier
- ! other real-time PCR devices
- ! appropriate reaction volumes, other than 50 µl

has to be validated in the lab by the user. The special notes regarding the internal control IC have to be strongly followed.

2. KIT CONTENTS

Universal Positive Control Sample (PC)	1 vial, 1 ml
Ready Master Mix (RMM), lyophilized	96 test-tubes 12 strips x 8 tubes
The kit is additionally supplied with optical-quality PCR-film	

3. PRINCIPLE OF THE METHOD

The Real time PCR is based on the detection of the fluorescence, produced by a reporter molecule, which increases as the reaction proceeds. Reporter molecule is dual-labeled DNA-probe, which specifically binds to the target region of pathogen DNA. Fluorescent signal increases due to the fluorescent dye and quencher separating by Taq DNA-polymerase exonuclease activity during amplification. PCR process consists of repeated cycles: temperature denaturation of DNA, primer annealing and complementary chain synthesis.

Threshold cycle value – Ct – is the cycle number at which the fluorescence generated within a reaction crosses the fluorescence threshold, a fluorescent signal rises significantly above the background fluorescence. Ct depends on initial quantity of pathogen DNA template.

The use of **Internal Control (IC)** prevents generation of false negative results associated with possible loss of DNA template during specimen preparation. IC indicates if PCR inhibitors occur in the reaction mixture. IC template should be added in each single sample (including control samples) prior to DNA extraction procedure. The amplification and detection of IC does not influence the sensitivity or specificity of the target DNA PCR.

Note: IC is a component of the NA extraction kits of RealLine series. Internal Control is added to the sample during NA isolation step and is used throughout the whole process of NA extraction, amplification, detection.

RealLine HSV-1 / HSV-2 Str-Format

4. SPECIFICATIONS

4.1. Specificity of HSV-1/HSV-2 DNA detection was determined using the Standard Reference Panel of negative DNA-extracts, consisting of samples containing IC DNA and not containing DNA of STD agents. Specificity of HSV-1/HSV-2 DNA detection equals 100%.

4.2. Sensitivity control was performed on 5 samples containing 100 HSV-1/HSV-2 DNA copies per sample, prepared from Standard Reference Samples containing DNA of HSV 1 and 2 types, (HSV-1/HSV-2 DNA SRS. The sensitivity equals 100%.

4.3. Diagnostic evaluation:

Diagnostic evaluation was performed on 50 clinical samples:

20 samples, negative samples;

20 samples, obtained from HSV-1/HSV-2 infected individuals;

10 samples, obtained from individuals infected with other types of HSV.

Determination of sensitivity was performed on 20 clinical samples obtained from the clinical material containing HSV types 1 and 2 with the CE-marked reference kit. **RealLine HSV-1 / HSV-2** kit determined all 20 samples as positive. Analysis by the reference kit proved all 20 samples containing HSV-1/HSV-2 to be positive.

Diagnostic sensitivity equals 100%.

Determination of specificity was performed on 30 samples obtained from donors and individuals infected with other types of HSV, with the CE-marked reference kit.

When studying clinical samples obtained from healthy donors **RealLine HSV-1 / HSV-2**, negative results were recorded for all 20 samples.

When studying clinical samples obtained from individuals infected by other types of HSVs using **RealLine HSV-1 / HSV-2** negative results were recorded for all 10 samples.

Specificity equals 100%.

Analysis by the CE-marked reference kit showed full match of results.

RealLine HSV-1 / HSV-2 Str-Format

5. PRODUCT USE LIMITATIONS

- This assay must not be used on the clinical specimen directly. Appropriate nucleic acids extraction methods have to be conducted prior to using this assay.
- The presence of PCR inhibitors (e.g. heparin) may cause false negative or invalid results.
- When monitoring a patient the same extraction method must be used in all determinations. Otherwise, results may not be comparable.
- The kit is designed for use in patients with a clinical history and/or symptoms consistent with HSV infections. The kit may be used for screening purposes.
- Diagnostic sensitivity of the kit may vary depending on the pathogen prevalence and characteristics of the enrolled cohort.
- Reliable results depend on adequate specimen sampling.
- Positive results indicate active or asymptomatic infection; clinical history and symptoms should be taken into account.
- Negative results indicate lack of detectable DNA but do not exclude the infection or disease.
- Potential mutations within the target regions of the HSV-1, respectively HSV-2 genome covered by the primers and/or probes used in the kit may result in failure to detect the presence of the pathogens.
- The kit is not intended to replace culture and other methods (e.g., cervical exam) for diagnosis of infections.

RealLine HSV-1 / HSV-2 Str-Format

6. WARNING AND PRECAUTIONS

- ☞ For In vitro use only.
- ☞ The kits must be used by skilled personnel only.
- ☞ When handling the kit, follow the national safety requirements for working with pathogens.
- ☞ To prevent contamination, the stages of DNA isolation and PCR test run must be spatially separated.
- ☞ Avoid microbial and nuclease contamination of reagents when removing aliquots from reagent vials.
- ☞ Wear protective disposable gloves, laboratory coats and eye protection when handling specimens and kit reagents.
- ☞ Every workplace must be provided with its own set of variable-volume pipettes, necessary auxiliary materials and equipment. It is prohibited to relocate them to other workplaces.
- ☞ The use of sterile disposable pipette tips is recommended.
- ☞ Never use the same tips for different samples.
- ☞ Do not pool reagents from different lots or from different vials of the same lot.
- ☞ Dispose unused reagents and waste in accordance with country, federal, state and local regulations.
- ☞ Do not use the kit after the expiration date at the side label of the kit.

7. ADDITIONAL MATERIALS AND DEVICES REQUIRED BUT NOT SUPPLIED

- real time PCR system, like described in p.1
- DNA-Extraction Kit, RealLine DNA-Extraction 3 or see p.1 Extractions Kits with Internal control reagent;
- Internal Control reagent (VBC8881) and Negative Control Sample if the kit is used with the extraction kits of other supplier;
- laminar safety box;
- refrigerator;
- half-automatic variable-volume single-channel pipettes;
- disposable medical non-sterile powder-free gloves;
- disposable pipette tips with aerosol barrier;
- biohazard waste container;
- razor or scalpel.

RealLine HSV-1 / HSV-2 Str-Format

8. PREPARATION OF THE SPECIMEN

The assay is performed on extracted DNA samples obtained from the clinical material using one of the DNA extraction kits listed in p.1, according to the Instruction Manual to the kit.

*Each group of samples undergoing the procedure of DNA isolation must include a **Positive Control sample (PC)** from this kit and a **Negative Control sample (NC)** which is a component of the DNA extraction kit.*

We strongly recommend the implementation of the Internal Control IC, the Negative Control NC and Positive Control PC samples to the extraction procedure.

When using kits of another supplier for the extraction of nucleic acids as recommended in chapter 1: add **20 µl** of **IC (VBC8881)** to each tube.

- For the NC use **100 µl** of the Negative Control Sample
- For the PC use **70 µl** of Negative Control Sample and **30 µl** of Positive Control to the tube marked PC.

If samples of isolated DNA were stored frozen prior the assay, thaw them and keep at least 30 minutes at a temperature of (18 – 25) °C.

The isolated DNA can be stored at (2 – 8) °C for 2 days.

After initial opening shelf life of Positive Control sample is 1 month at (2 – 8) °C or for 50 µl aliquots 3 month at (-18 ... -60) °C

RealLine HSV-1 / HSV-2 Str-Format

9. PROCEDURE

9.1. Preparation of the reagents.

Prior to the test take the kit out of the refrigerator and keep the **Ready Master Mix (RMM)** closed in the package at (18 – 25) °C for at least 30 minutes. Then open the package and cut the necessary number of tubes in strips with RMM (including prepared samples and controls: 1 NC and 1 PC) with the razor. Cut the tubes together with the covering film..

Put the remaining strips immediately back into the foil pouch, squeeze the air out and tightly close with the clip.

After initial opening store RMM at (2 – 8) °C for no more than 3 months.

9.2. Label the tubes with RMM for each specimen and control.

Attention! Labels should be placed on the lateral side of the tubes.

9.3. Add **50 µl** of corresponding isolated DNA solution to each tube using a separate pipette tip with filter. Do not touch the pellet! Tightly close the tubes with caps or seal with the PCR transparent film.

9.4. Place the tubes into the Real Time PCR system.

Program Real Time PCR system as follows:

Step 1:	50°C	2min	
Step 2:	95°C	2min	
Step 3:	94°C	10 sec	50 Cycles
	60°C*	20 sec	

* Measurement of fluorescence at 60°C in channels FAM, HEX, and ROX

9.5. Select the amplification detection channels:

- Collect Real-time PCR data through the **FAM** channel for detection of amplification of **IC DNA**.
- Collect Real-time PCR data through the **HEX** channel for detection of amplification of **HSV-1 DNA**.
- Collect Real-time PCR data through the **ROX** channel for detection of amplification of **HSV-2 DNA**.
-

9.6. Program the positions of test tubes with samples, positive and negative controls according to the instruction manual for the Real Time PCR system in use.

9.7. Run the program.

RealLine HSV-1 / HSV-2 Str-Format

10. DATA ANALYSIS AND INTERPRETATION

10.1 For **PC** the program should detect:

- increase of the **IC** DNA amplification signal (channel **FAM**) and determine the threshold cycle, **IC Ct**;
- increase of the **HSV-1** DNA amplification signal (channel **HEX**) and determine the **Ct** value;
- increase of the **HSV-2** DNA amplification signal (channel **ROX**) and determine the **Ct** value.

10.2 For **NC** the program should detect the increase of the amplification signal of IC DNA (channel **FAM**) and determine the threshold cycle, **IC Ct**. No significant **HEX** and **ROX** fluorescent increase should appear (*no HSV-1 and HSV-2 DNA amplification*).

10.3 For each specimen the program should detect the increase of the amplification signal of IC DNA (channel **FAM**) and determine **IC Ct**.

10.4 Calculate $(IC\ Ct)_{av}$ as an average **IC Ct** of all analyzed samples (including PC and NC). **IC Ct** values that differ by more than 2 cycles from the $(IC\ Ct)_{av}$ should be ignored. Recalculate the $(IC\ Ct)_{av}$ for the remaining values after the screening.

10.5 The sample is considered **positive** (containing HSV-1 DNA) when **Ct** value via **HEX** channel for this sample is **less than or equals to 40**.

The sample is considered **positive** (containing HSV-2 DNA) when **Ct** value via **ROX** channel for this sample is **less than or equals to 40**.

10.6 The sample is considered **negative** (not containing HSV-1 or HSV-2 DNA), if **Ct** value via **HEX** and **ROX** channels for this sample is **above 40** or is not determined.

If **IC Ct** value for such sample differs from the $(IC\ Ct)_{av}$ value by more than 2 cycles, the result is regarded as equivocal. A repeated analysis of the sample, starting with the DNA isolation step is necessary.

10.7 If the **Ct** value for NC through the “**ROX**” or “**HEX**” channel is **less than or equal to 40**, this indicates the presence of contamination. In case of contamination all positive results of this individual PCR test run are considered as equivocal. Actions are required to identify and eliminate the source of contamination, and repeat the analysis of all samples of this run that were identified as positive. Samples that showed negative results in this run should be considered negative.

RealLine HSV-1 / HSV-2 Str-Format

11. STORAGE AND TRANSPORTATION

- Store and transport the assay kit at (2 - 8) °C in the manufacturer's packing.
- Transportation at 25 °C for up to 10 days is allowed.
- Do not freeze the kit!
- Do not pool reagents from different lots or from different vials of the same lot.
- Strictly follow the Instruction manual for reliable results.
- Do not use kits with damaged inner packages and get in contact with BIORON Diagnostics GmbH.

- **Storage and shelf life of solutions and components of the kit after initial opening:**
 - Positive Control sample: 1 month at (2 - 8) °C.
 - Ready Master Mix (RMM): 3 months at (2 - 8) °C

Technical Support: techsupport@bioron.de

ANNEX I: Settings for RealLine Cyclers and DT96:

for these cyclers the measurement exposure must be adjusted. Choose the **Operation with the device** mode in the **Settings** menu, select the item **Measurement exposition:**

- **FAM** to **250**
- **HEX** and **ROX** to **1000**

Confirm that the current exposure value is saved by pressing **YES**

Attention! The specified exposure values are applicable only for RealLine kits and, if necessary, must be changed for other purposes.

